



**Republic of the Philippines**  
**Department of Education**  
 Region I  
 Schools Division Office  
 San Carlos City, Pangasinan

Reference No: 20240802-001

**SUPPLEMENTAL/BID BULLETIN NO.: 11084520-1;**

**SUPPLY AND DELIVERY OF DISASTER AND EMERGENCY- RELATED EQUIPMENT  
 FOR THE IMPLEMENTATION OF DRRM PPA's UNDER FY 2024 DISASTER  
 PREPAREDNESS AND RESPONSE PROGRAM (DPRP)**

**ABC: PHP 1,386,000.00**

The Schools Division Office of San Carlos City Bids and Awards Committee would like to inform you of the following additional instructions and amendments made on the Bidding Documents for the abovementioned project, to wit:

<b>SUPPLY AND DELIVERY OF DISASTER AND EMERGENCY- RELATED EQUIPMENT            FOR THE IMPLEMENTATION OF DRRM PPA's UNDER FY 2024 DISASTER            PREPAREDNESS AND RESPONSE PROGRAM (DPRP)</b>	
FROM	TO
<p><b>1. SECTION VII. TECHNICAL SPECIFICATIONS</b> PAGE 35, ITEM DESCRIPTION, ITEM NO. 3</p> <p>WHEELCHAIR</p> <p>-Seat width: 18"/20"</p> <ul style="list-style-type: none"> <li>- Frame color: painted black covered frame</li> <li>- Fixed PVC padded armrest with black color</li> <li>- Fixed foot rest</li> <li>- Stainless steel side panel</li> <li>- 8" front castor with solid tire</li> <li>- 24" rear spoke and aluminum wheel with solid tire</li> <li>- Steel hand rim</li> <li>- Plastic footplate</li> <li>- PVC cushion</li> <li>- Foldable</li> </ul>	<p><b>SECTION VII. TECHNICAL SPECIFICATIONS</b> PAGE 35, ITEM DESCRIPTION, ITEM NO. 3</p> <p>WHEELCHAIR</p> <p>-Seat width: <b>18"- 20"</b> <b>Rotation: 360° front wheel</b></p> <ul style="list-style-type: none"> <li>- Frame color: painted black covered frame <b>(Manufacturer standard)</b></li> <li>- <b>Foldable</b> foot rest</li> <li>-Stainless steel</li> <li>- 8" front castor with solid tire</li> <li>- 24" rear spoke and aluminum wheel with solid tire</li> <li>- Steel hand rim</li> <li>- Plastic footplate</li> <li>- PVC cushion</li> <li>- Foldable</li> </ul>
<p><b>2. SECTION VII. TECHNICAL SPECIFICATIONS</b> PAGE 35, ITEM DESCRIPTION, ITEM NO. 4</p> <p>DIGITAL BLOOD PRESSURE</p>	<p><b>SECTION VII. TECHNICAL SPECIFICATIONS</b> PAGE 35, ITEM DESCRIPTION, ITEM NO. 4</p> <p>DIGITAL BLOOD PRESSURE</p>



Address: Roxas Blvd., San Carlos City, Pangasinan  
 Tel. No./Fax No.: (075) 632-3293  
 Email Address: [sdosancarloscity@gmail.com](mailto:sdosancarloscity@gmail.com)  
[sancarlos.city1@deped.gov.ph](mailto:sancarlos.city1@deped.gov.ph)



Republic of the Philippines  
**Department of Education**

Region I  
 Schools Division Office  
 San Carlos City, Pangasinan

<p>*Not included</p> <ul style="list-style-type: none"> <li>- Measuring range: Pressure: 0mmHg - 280mmhg Pulse: (40 - 199)/mi</li> <li>- Accuracy/calibration: Pressure: <math>\pm .4kPa(\pm 3mmHg)</math> Pules: <math>\pm 5\%</math></li> <li>- Operating temperature/Humidity: 5'C - 40'C / 10%RH - 90%RH</li> <li>- Storage Temperature: -20'C - 65'C /10%RH - 95%RH</li> <li>- Working Principle: Oscillometer</li> <li>- Display: LCD</li> <li>- Size: L140mm*W100mm*H80mm</li> <li>- Weight: About 430 g</li> <li>- Power: 1.5V (Alkaline.AA) x4</li> <li>- With 1-year warranty</li> </ul>	<p><b>-Two-way power supply (battery and power port)</b></p> <p>Measuring range: Pressure: 0mmHg - 280mmhg Pulse: (40 - 199)/mi</p> <ul style="list-style-type: none"> <li>- Accuracy/calibration: Pressure: <math>\pm .4kPa(\pm 3mmHg)</math> Pules: <math>\pm 5\%</math></li> <li>- Operating temperature/Humidity: 5'C - 40'C / 10%RH - 90%RH</li> <li>- Storage Temperature: -20'C - 65'C /10%RH - 95%RH</li> <li>- Working Principle: Oscillometer</li> <li>- Display: LCD</li> <li>- Size: L140mm*W100mm*H80mm</li> <li>- Weight: About 430 g</li> <li>- Power: 1.5V (Alkaline.AA) x4</li> <li>- With 1-year warranty</li> </ul>
<p><b>SECTION VII. TECHNICAL SPECIFICATIONS</b>  <b>PAGE 36, ITEM DESCRIPTION, ITEM NO. 5</b></p> <p>PULSE OXIMETER          *Not included          Display Information:</p> <ul style="list-style-type: none"> <li>- The Pulse Oxygen Saturation (SpO2) - OLED</li> <li>- Pulse Rate (PR) - OLED</li> <li>- Pulse Intensity (bar-graph) - OLED bar-graph display</li> <li>- Pulse Wave - OLED</li> </ul> <p>SpO2 Parameter Specification:</p> <ul style="list-style-type: none"> <li>- Measuring range: 0%-100% (the resolution is 1%)</li> <li>- Accuracy: 70%-100%<math>\pm 2\%</math>; Below 70% unspecified</li> <li>- Optical Sensor: Red light (wavelength is 660nm); Infrared (wavelength is 880nm) with 1-year warranty</li> </ul> <p>Pulse Parameter Specification:</p> <ul style="list-style-type: none"> <li>- Measuring range: 30bpm-250bpm (the resolution is 1 bpm)</li> <li>- Accuracy: <math>\pm 2bpm</math> or <math>\pm 2\%</math> select larger</li> </ul> <p>Pulse Intensity:</p>	<p><b>SECTION VII. TECHNICAL SPECIFICATIONS</b>  <b>PAGE 36, ITEM DESCRIPTION, ITEM NO. 5</b></p> <p>PULSE OXIMETER  <b>- Finger tip</b>          Display Information:</p> <ul style="list-style-type: none"> <li>- The Pulse Oxygen Saturation (SpO2) - OLED</li> <li>- Pulse Rate (PR) - OLED</li> <li>- Pulse Intensity (bar-graph) - OLED bar-graph display</li> <li>- Pulse Wave - OLED</li> </ul> <p>SpO2 Parameter Specification:</p> <ul style="list-style-type: none"> <li>- Measuring range: 0%-100% (the resolution is 1%)</li> <li>- Accuracy: 70%-100%<math>\pm 2\%</math>; Below 70% unspecified</li> <li>- Optical Sensor: Red light (wavelength is 660nm); Infrared (wavelength is 880nm) with 1-year warranty</li> </ul> <p>Pulse Parameter Specification:</p> <ul style="list-style-type: none"> <li>- Measuring range: 30bpm-250bpm (the resolution is 1 bpm)</li> <li>- Accuracy: <math>\pm 2bpm</math> or <math>\pm 2\%</math> select larger</li> </ul> <p>Pulse Intensity:</p>






Republic of the Philippines  
**Department of Education**  
Region I  
Schools Division Office  
San Carlos City, Pangasinan

<ul style="list-style-type: none"><li>- Range: Continuous bar-graph display, the higher display indicates the stronger pulse</li></ul> Battery Useful Life: <ul style="list-style-type: none"><li>- Battery can work continually for 20 hours (approximately), please charge once it shows low battery</li></ul> Dimensions and Weight: <ul style="list-style-type: none"><li>- Dimensions: 57(L) x 31(w) x 32 (H) mm</li><li>- Weight: About 50g (with the batteries)</li><li>- With 1-year warranty</li></ul>	<ul style="list-style-type: none"><li>- Range: Continuous bar-graph display, the higher display indicates the stronger pulse</li></ul> Battery Useful Life: <ul style="list-style-type: none"><li>- Battery can work continually for 20 hours (approximately), please charge once it shows low battery</li></ul> Dimensions and Weight: <ul style="list-style-type: none"><li>- Dimensions: 57(L) x 31(w) x 32 (H) mm</li><li>- Weight: About 50g (with the batteries)</li><li>- With 1-year warranty</li></ul>
---	---

All other provisions of the bidding documents which are not affected shall remain in force and in effect.

Please be guided accordingly.

  
**DOMINGO L. LAUD, EdD, CESE**  
OIC, Assistant Schools Division Superintendent  
Chairperson, Bids and Awards Committee

